K012568

SEP - 6 2001

Attachment B
Summary of Safety and Effectiveness

Prepared in accordance with 21 CFR Part 807.92(c).

GE Medical Systems

General Electric Company
8380 Darrow Road, Twinsburg, OH 44087

Submitter: GE Medical Systems - SMV
8380 Darrow Road
Twinsburg, OH 44087

Contact Person: Kevin Murrock
Engineering Technical Services
Telephone: 330-487-6635; Fax: 330-405-7684

Date Prepared: 31-July-2001

Device Name: GE Vision Nuclear Medicine Workstation.
Emission Computed Tomography System, 21 CFR 892.1200, 90-KPS

Marketed Device: GE Vision POWERstation Nuclear Medicine Workstation, 510(k) Number K912573, currently in commercial distribution.

Device Description: The GE Vision Nuclear Medicine Workstation is a nuclear medicine image display and processing workstation. It consists of a personal computer (PC) or UNIX workstation, monitor, keyboard, mouse, modem, and network interface.

Indications for Use: The GE Vision Nuclear Medicine Workstation is intended for use in the display and analysis of planar and tomographic nuclear medicine images.

Comparison with Predicate Device: The GE Vision Nuclear Medicine Workstation is of a comparable type and substantially equivalent to the currently marketed GE Vision POWERstation Nuclear Medicine Workstation. It has the same technological characteristics, is comparable in key safety and effectiveness features, uses the same basic design, and has the same intended uses as the predicate device.

Summary of Studies: The device has been evaluated for electrical, mechanical, and radiation safety, and conforms with applicable medical device safety standards.

Clinical Tests: None required.

Conclusion: Intended uses and other key features are consistent with traditional clinical practice, FDA guidelines, and established methods of patient examination. Intended uses and fundamental scientific technology are the same as the legally marketed GE Vision POWERstation Nuclear Medicine Workstation. The design and development process of the manufacturer conforms with 21 CFR 820, ISO 9001 and EN 46001 quality systems. The device conforms to applicable medical device safety standards and compliance is verified through independent evaluation with factory surveillance. Therefore, it is the opinion of GE Medical Systems that the GE Vision Nuclear Medicine Workstation is substantially equivalent with respect to safety and effectiveness to devices currently cleared for market.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

SEP - 6 2001

Mr. Kevin M. Murrock
Manager, Engineering Technical Services
GE Medical Systems
General Electric Company
8380 Darrow Road
TWINSBURG OH 44087

Re: K012568
GE Vision Nuclear Medicine Workstation
(Emission Computed Tomography System)
Dated: July 31, 2001
Received: August 9, 2001
Regulatory Class: II
21 CFR 892.1200/Procode: 90 KPS

Dear Mr. Murrock:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at one of the following numbers, based on the regulation number at the top of this letter:

8xx.1xxx	(301) 594-4591
876.2xxx, 3xxx, 4xxx, 5xxx	(301) 594-4616
884.2xxx, 3xxx, 4xxx, 5xxx, 6xxx	(301) 594-4616
892.2xxx, 3xxx, 4,xxx, 5xxx	(301) 594-4654
Other	(301) 594-4692

Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/dsma/dsmamain.html.

Sincerely yours,

Nancy C. Brogdon

Nancy C. Brogdon
Director, Division of Reproductive,
Abdominal, and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

STATEMENT OF INTENDED USE

510(k) Number (if known): K012568

Device Name: GE Vision Nuclear Medicine Workstation

Indications for Use

The GE Vision Nuclear Medicine Workstation is intended for use in the display and analysis of planar and tomographic nuclear medicine images.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓
(Per 21 CFR 801-109)

OR Over-The-Counter Use_____

Nancy C Brogdon

(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number K012568